

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100918\
 Data File : VX005197.D
 Acq On : 10 Oct 2018 09:13
 Operator : JC/MD
 Sample : J5303-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW515-3034

Quant Time: Oct 10 16:32:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	180335	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	286634	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	286937	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	161089	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	141033	54.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.50%	
35) Dibromofluoromethane	5.50	113	119671	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
50) Toluene-d8	8.72	98	416650	51.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.18%	
62) 4-Bromofluorobenzene	11.14	95	150698	51.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.60%	
Target Compounds						
16) Acetone	2.45	43	4298	2.858	ug/l	93
24) 1,1-Dichloroethane	3.70	63	4706	1.099	ug/l #	94
64) Tetrachloroethene	9.34	164	5532	1.959	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100918\
Data File : VX005197.D
Acq On : 10 Oct 2018 09:13
Operator : JC/MD
Sample : J5303-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW515-3034

Quant Time: Oct 10 16:32:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

