

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100918\
 Data File : VX005200.D
 Acq On : 10 Oct 2018 10:33
 Operator : JC/MD
 Sample : J5303-17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW516-2024

Quant Time: Oct 10 16:35:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	175768	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	278637	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	279706	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	156867	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	144244	57.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.90%	
35) Dibromofluoromethane	5.50	113	117311	49.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.36%	
50) Toluene-d8	8.72	98	405090	51.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.20%	
62) 4-Bromofluorobenzene	11.14	95	145849	51.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.14%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100918\
Data File : VX005200.D
Acq On : 10 Oct 2018 10:33
Operator : JC/MD
Sample : J5303-17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW516-2024

Quant Time: Oct 10 16:35:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

