

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100918\
 Data File : VX005201.D
 Acq On : 10 Oct 2018 11:00
 Operator : JC/MD
 Sample : J5303-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW516-3034

Quant Time: Oct 19 16:09:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	168587	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	273450	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	269360	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	151582	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	136080	56.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.02%	
35) Dibromofluoromethane	5.51	113	111739	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	
50) Toluene-d8	8.72	98	394189	51.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.32%	
62) 4-Bromofluorobenzene	11.14	95	140804	50.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.46%	
Target Compounds						
16) Acetone	2.45	43	4491	3.194	ug/l	99
24) 1,1-Dichloroethane	3.70	63	12201	3.048	ug/l	98
64) Tetrachloroethene	9.34	164	10943	4.128	ug/l	96
85) sec-Butylbenzene	11.94	105	5068	0.529	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100918\
Data File : VX005201.D
Acq On : 10 Oct 2018 11:00
Operator : JC/MD
Sample : J5303-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW516-3034

Quant Time: Oct 19 16:09:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

