

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100918\
 Data File : VX005202.D
 Acq On : 10 Oct 2018 11:27
 Operator : JC/MD
 Sample : J5303-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW516-4044

Quant Time: Oct 10 16:40:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	171279	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	270698	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	269960	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	150535	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	135403	55.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.68%	
35) Dibromofluoromethane	5.50	113	114184	49.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.54%	
50) Toluene-d8	8.72	98	393875	51.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.28%	
62) 4-Bromofluorobenzene	11.14	95	141191	51.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.78%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	982	0.524	ug/l #	90
16) Acetone	2.45	43	3914	2.740	ug/l	98
24) 1,1-Dichloroethane	3.70	63	6337	1.558	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100918\
Data File : VX005202.D
Acq On : 10 Oct 2018 11:27
Operator : JC/MD
Sample : J5303-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW516-4044

Quant Time: Oct 10 16:40:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

