

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100919\
 Data File : VX012934.D
 Acq On : 09 Oct 2019 11:53
 Operator : JC/SP
 Sample : K5214-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Oct 10 02:32:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	188948	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	317473	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	268077	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	103377	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	122551	51.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.78%	
35) Dibromofluoromethane	5.49	113	95162	52.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.46%	
50) Toluene-d8	8.71	98	372470	52.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.82%	
62) 4-Bromofluorobenzene	11.13	95	125487	45.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.88%	
Target Compounds						
16) Acetone	2.44	43	4212	3.384	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100919\
Data File : VX012934.D
Acq On : 09 Oct 2019 11:53
Operator : JC/SP
Sample : K5214-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Oct 10 02:32:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

