

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100919\
 Data File : VX012938.D
 Acq On : 09 Oct 2019 13:26
 Operator : JC/SP
 Sample : PB123722TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB123722TB

Quant Time: Oct 10 08:38:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	182688	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	311079	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	272950	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	106826	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	119313	52.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.50%	
35) Dibromofluoromethane	5.49	113	91145	51.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.10%	
50) Toluene-d8	8.71	98	363562	52.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.42%	
62) 4-Bromofluorobenzene	11.14	95	125733	46.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.94%	
Target Compounds						
20) Methylene Chloride	2.85	84	3938	1.780	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100919\
Data File : VX012938.D
Acq On : 09 Oct 2019 13:26
Operator : JC/SP
Sample : PB123722TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB123722TB

Quant Time: Oct 10 08:38:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

