

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100919\
 Data File : VX012939.D
 Acq On : 09 Oct 2019 13:49
 Operator : JC/SP
 Sample : K5234-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-MMW0008

Quant Time: Oct 10 03:49:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	174111	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	298562	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	247266	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	91176	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	115325	52.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.98%	
35) Dibromofluoromethane	5.49	113	87862	51.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.54%	
50) Toluene-d8	8.71	98	343397	51.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.76%	
62) 4-Bromofluorobenzene	11.14	95	113143	43.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.14%	
Target Compounds						
16) Acetone	2.43	43	2691	2.346	ug/l #	85
37) Ethyl Acetate	4.83	43	12634	4.304	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100919\
Data File : VX012939.D
Acq On : 09 Oct 2019 13:49
Operator : JC/SP
Sample : K5234-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-MMW0008

Quant Time: Oct 10 03:49:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

