

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100919\
 Data File : VX012941.D
 Acq On : 09 Oct 2019 14:36
 Operator : JC/SP
 Sample : K5234-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-JBW7106A

Quant Time: Oct 10 08:44:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	176982	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	299000	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	252986	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	93825	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	116692	52.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.50%	
35) Dibromofluoromethane	5.49	113	89173	51.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.92%	
50) Toluene-d8	8.71	98	349522	52.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.44%	
62) 4-Bromofluorobenzene	11.14	95	117487	45.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.36%	
Target Compounds						
7) Trichlorofluoromethane	1.92	101	1593	0.582	ug/l	85
16) Acetone	2.44	43	6607	5.667	ug/l #	84
37) Ethyl Acetate	4.83	43	37726	12.834	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100919\
Data File : VX012941.D
Acq On : 09 Oct 2019 14:36
Operator : JC/SP
Sample : K5234-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-JBW7106A

Quant Time: Oct 10 08:44:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

