

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100919\
 Data File : VX012942.D
 Acq On : 09 Oct 2019 14:59
 Operator : JC/SP
 Sample : K5234-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-MMW0127

Quant Time: Oct 10 03:57:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	174476	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	291521	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	244256	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	90738	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	113812	52.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.38%	
35) Dibromofluoromethane	5.49	113	87585	52.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.70%	
50) Toluene-d8	8.71	98	340092	52.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.24%	
62) 4-Bromofluorobenzene	11.14	95	113613	44.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.62%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100919\
Data File : VX012942.D
Acq On : 09 Oct 2019 14:59
Operator : JC/SP
Sample : K5234-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-MMW0127

Quant Time: Oct 10 03:57:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

