

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100919\
 Data File : VX012944.D
 Acq On : 09 Oct 2019 15:46
 Operator : JC/SP
 Sample : K5234-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-MMW0019

Quant Time: Oct 10 08:51:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	179018	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	301841	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	258100	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	98737	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	117549	52.54	ug/l	0.00
Spiked Amount			Recovery	=	105.08%	
35) Dibromofluoromethane	5.49	113	90512	52.25	ug/l	0.00
Spiked Amount			Recovery	=	104.50%	
50) Toluene-d8	8.71	98	351546	52.03	ug/l	0.00
Spiked Amount			Recovery	=	104.06%	
62) 4-Bromofluorobenzene	11.14	95	118188	45.02	ug/l	0.00
Spiked Amount			Recovery	=	90.04%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100919\
Data File : VX012944.D
Acq On : 09 Oct 2019 15:46
Operator : JC/SP
Sample : K5234-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-MMW0019

Quant Time: Oct 10 08:51:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

