

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100919\
 Data File : VX012958.D
 Acq On : 09 Oct 2019 21:57
 Operator : JC/SP
 Sample : VX1009WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1009WBL02

Quant Time: Oct 10 06:59:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	176130	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	297112	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	252390	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	93980	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	114582	52.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.10%	
35) Dibromofluoromethane	5.49	113	88807	52.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.16%	
50) Toluene-d8	8.71	98	348846	52.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.90%	
62) 4-Bromofluorobenzene	11.14	95	115134	44.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.10%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100919\
Data File : VX012958.D
Acq On : 09 Oct 2019 21:57
Operator : JC/SP
Sample : VX1009WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1009WBL02

Quant Time: Oct 10 06:59:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

