

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100919\  
 Data File : VX012961.D  
 Acq On : 09 Oct 2019 23:30  
 Operator : JC/SP  
 Sample : K5260-02  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 37 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 FA19-JBW7203A

Quant Time: Oct 10 07:11:04 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X100819W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Oct 09 01:51:23 2019  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	182489	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	304186	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	264391	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	105171	50.00	ug/l	0.00

## System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.05	65	118312	51.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.74%	
35) Dibromofluoromethane	5.49	113	90220	51.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.36%	
50) Toluene-d8	8.71	98	358414	52.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.28%	
62) 4-Bromofluorobenzene	11.14	95	125130	47.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.58%	

## Target Compounds

					Qvalue
3) Chloromethane	1.32	50	2566	1.268	ug/l 99
11) Tert butyl alcohol	3.02	59	2591	4.285	ug/l # 93
16) Acetone	2.43	43	13748	11.435	ug/l 93
37) Ethyl Acetate	4.83	43	11153	3.729	ug/l # 84
44) Trichloroethene	7.20	130	2859	1.305	ug/l 82
64) Tetrachloroethene	9.33	164	10903	5.946	ug/l 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100919\  
Data File : VX012961.D  
Acq On : 09 Oct 2019 23:30  
Operator : JC/SP  
Sample : K5260-02  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 37 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
FA19-JBW7203A

Quant Time: Oct 10 07:11:04 2019  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X100819W.M  
Quant Title : SW846 8260  
QLast Update : Wed Oct 09 01:51:23 2019  
Response via : Initial Calibration

