

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100919\
 Data File : VX012962.D
 Acq On : 09 Oct 2019 23:54
 Operator : JC/SP
 Sample : K5260-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-MMW0010

Quant Time: Oct 10 07:16:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	184186	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	308136	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	262292	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	103800	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	117672	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
35) Dibromofluoromethane	5.49	113	90990	51.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.90%	
50) Toluene-d8	8.71	98	358063	51.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.82%	
62) 4-Bromofluorobenzene	11.14	95	123298	46.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.00%	
Target Compounds						
16) Acetone	2.44	43	8981	7.401	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	1517	0.660	ug/l	97
44) Trichloroethene	7.20	130	3246	1.463	ug/l	98
64) Tetrachloroethene	9.33	164	6388	3.511	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100919\
Data File : VX012962.D
Acq On : 09 Oct 2019 23:54
Operator : JC/SP
Sample : K5260-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-MMW0010

Quant Time: Oct 10 07:16:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

