

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100919\
 Data File : VX012964.D
 Acq On : 10 Oct 2019 00:40
 Operator : JC/SP
 Sample : K5260-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-MMW0011

Quant Time: Oct 10 07:19:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	178223	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	299399	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	256907	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	96600	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	116103	52.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.24%	
35) Dibromofluoromethane	5.49	113	88339	51.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.82%	
50) Toluene-d8	8.71	98	349833	52.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.40%	
62) 4-Bromofluorobenzene	11.14	95	117758	45.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.44%	
Target Compounds						
7) Trichlorofluoromethane	1.93	101	1081	0.392	ug/l	91
16) Acetone	2.43	43	12877	10.967	ug/l	99
27) cis-1,2-Dichloroethene	4.59	96	1497	0.673	ug/l	93
44) Trichloroethene	7.21	130	3336	1.547	ug/l	84
64) Tetrachloroethene	9.33	164	6331	3.553	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100919\
Data File : VX012964.D
Acq On : 10 Oct 2019 00:40
Operator : JC/SP
Sample : K5260-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-MMW0011

Quant Time: Oct 10 07:19:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

