

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100919\
 Data File : VX012966.D
 Acq On : 10 Oct 2019 01:27
 Operator : JC/SP
 Sample : K5260-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-MMW0013

Quant Time: Oct 10 07:30:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	178731	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	303723	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	259577	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	102378	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	116750	52.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.52%	
35) Dibromofluoromethane	5.49	113	90029	51.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.30%	
50) Toluene-d8	8.71	98	353371	51.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.94%	
62) 4-Bromofluorobenzene	11.14	95	119811	45.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.70%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.32	50	1206	0.608	ug/l #	85
7) Trichlorofluoromethane	1.93	101	1044	0.378	ug/l	96
11) Tert butyl alcohol	3.03	59	4717	7.965	ug/l #	94
16) Acetone	2.43	43	12939	10.989	ug/l	95
27) cis-1,2-Dichloroethene	4.58	96	1391	0.623	ug/l	89
37) Ethyl Acetate	4.82	43	44518	14.909	ug/l	98
44) Trichloroethene	7.22	130	7505	3.432	ug/l	92
64) Tetrachloroethene	9.33	164	10907	6.058	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100919\
Data File : VX012966.D
Acq On : 10 Oct 2019 01:27
Operator : JC/SP
Sample : K5260-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-MMW0013

Quant Time: Oct 10 07:30:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

