

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100919\
 Data File : VX012967.D
 Acq On : 10 Oct 2019 01:50
 Operator : JC/SP
 Sample : K5260-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-MMW0012

Quant Time: Oct 10 07:34:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	171507	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	286380	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	248578	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	95524	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	113142	52.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.56%	
35) Dibromofluoromethane	5.49	113	81571	49.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.26%	
50) Toluene-d8	8.71	98	344096	53.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.36%	
62) 4-Bromofluorobenzene	11.13	95	114614	46.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.02%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.02	59	5438	9.570	ug/l #	91
16) Acetone	2.43	43	31456	27.840	ug/l	93
25) 2-Butanone	4.68	43	4192	2.674	ug/l	92
27) cis-1,2-Dichloroethene	4.59	96	2840	1.326	ug/l	91
44) Trichloroethene	7.20	130	4317	2.093	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100919\
Data File : VX012967.D
Acq On : 10 Oct 2019 01:50
Operator : JC/SP
Sample : K5260-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-MMW0012

Quant Time: Oct 10 07:34:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

