

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100919\
 Data File : VX012968.D
 Acq On : 10 Oct 2019 02:13
 Operator : JC/SP
 Sample : K5260-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-JPZ7208

Quant Time: Oct 10 07:39:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	173781	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	290551	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	245685	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	90860	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	114143	52.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.10%	
35) Dibromofluoromethane	5.49	113	87711	52.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.20%	
50) Toluene-d8	8.71	98	341139	52.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.90%	
62) 4-Bromofluorobenzene	11.13	95	113117	44.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.52%	
Target Compounds						
11) Tert butyl alcohol	3.02	59	2863	4.972	ug/l #	77
16) Acetone	2.43	43	12418	10.847	ug/l	100
37) Ethyl Acetate	4.83	43	31010	10.856	ug/l	98
64) Tetrachloroethene	9.33	164	2611	1.532	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100919\
Data File : VX012968.D
Acq On : 10 Oct 2019 02:13
Operator : JC/SP
Sample : K5260-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-JPZ7208

Quant Time: Oct 10 07:39:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

