

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100919\
 Data File : VX012971.D
 Acq On : 10 Oct 2019 03:23
 Operator : JC/SP
 Sample : K5259-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-JBW7212B

Quant Time: Oct 10 07:44:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	174910	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	293154	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	251166	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	95841	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	114651	52.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.88%	
35) Dibromofluoromethane	5.49	113	88943	52.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.72%	
50) Toluene-d8	8.71	98	342009	52.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.24%	
62) 4-Bromofluorobenzene	11.14	95	117144	45.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.88%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	17381	7.959	ug/l	88
32) 1,1,1-Trichloroethane	5.48	97	1106	0.375	ug/l #	43
44) Trichloroethene	7.21	130	6147	2.912	ug/l	99
64) Tetrachloroethene	9.33	164	86893	49.879	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100919\
Data File : VX012971.D
Acq On : 10 Oct 2019 03:23
Operator : JC/SP
Sample : K5259-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-JBW7212B

Quant Time: Oct 10 07:44:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

