

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100919\
 Data File : VX012974.D
 Acq On : 10 Oct 2019 04:33
 Operator : JC/SP
 Sample : K5259-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-JMW1982

Quant Time: Oct 10 09:08:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	176454	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	296367	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	251236	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	97222	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	115802	52.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.02%	
35) Dibromofluoromethane	5.49	113	87189	51.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.52%	
50) Toluene-d8	8.71	98	346574	52.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.48%	
62) 4-Bromofluorobenzene	11.14	95	116511	45.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.40%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.58	96	16161	7.336	ug/l	88
44) Trichloroethene	7.21	130	25006	11.717	ug/l	94
64) Tetrachloroethene	9.33	164	57602	33.056	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100919\
Data File : VX012974.D
Acq On : 10 Oct 2019 04:33
Operator : JC/SP
Sample : K5259-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-JMW1982

Quant Time: Oct 10 09:08:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

