

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100919\
 Data File : VX012975.D
 Acq On : 10 Oct 2019 04:56
 Operator : JC/SP
 Sample : K5259-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-JMW1982-9001

Quant Time: Oct 10 07:52:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	174825	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	288620	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	241302	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	86324	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	113884	52.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.24%	
35) Dibromofluoromethane	5.49	113	87155	52.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.22%	
50) Toluene-d8	8.71	98	337279	52.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.40%	
62) 4-Bromofluorobenzene	11.14	95	109766	43.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.44%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	16834	7.713	ug/l	92
44) Trichloroethene	7.20	130	26812	12.901	ug/l	91
64) Tetrachloroethene	9.33	164	61528	36.762	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100919\
Data File : VX012975.D
Acq On : 10 Oct 2019 04:56
Operator : JC/SP
Sample : K5259-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-JMW1982-9001

Quant Time: Oct 10 07:52:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

