

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX100919\
 Data File : VX012929.D
 Acq On : 09 Oct 2019 09:33
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 10 01:56:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.510	0.550	-7.8	100	0.00
3 P	Chloromethane	0.555	0.548	1.3	96	0.00
4 C	Vinyl Chloride	0.573	0.579	-1.0#	98	0.00
5 T	Bromomethane	0.238	0.213	10.5	95	0.00
6 T	Chloroethane	0.337	0.348	-3.3	97	0.00
7 T	Trichlorofluoromethane	0.773	0.812	-5.0	101	0.00
8 T	Diethyl Ether	0.312	0.321	-2.9	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.500	0.537	-7.4	104	0.00
10 T	Methyl Iodide	0.492	0.533	-8.3	96	0.00
11 T	Tert butyl alcohol	0.166	0.136	18.1	83	0.00
12 CM	1,1-Dichloroethene	0.505	0.522	-3.4#	99	0.00
13 T	Acrolein	0.118	0.096	18.6	80	0.00
14 T	Allyl chloride	0.900	0.917	-1.9	96	0.00
15 T	Acrylonitrile	0.314	0.304	3.2	91	0.00
16 T	Acetone	0.329	0.388	-17.9	112	0.00
17 T	Carbon Disulfide	1.441	1.405	2.5	98	0.00
18 T	Methyl Acetate	0.781	0.725	7.2	91	0.00
19 T	Methyl tert-butyl Ether	1.745	1.814	-4.0	98	0.00
20 T	Methylene Chloride	0.605	0.586	3.1	97	0.00
21 T	trans-1,2-Dichloroethene	0.542	0.567	-4.6	102	0.00
22 T	Diisopropyl ether	1.724	1.804	-4.6	99	0.00
23 T	Vinyl Acetate	1.488	1.564	-5.1	96	0.00
24 P	1,1-Dichloroethane	0.970	1.008	-3.9	99	0.00
25 T	2-Butanone	0.457	0.480	-5.0	99	0.00
26 T	2,2-Dichloropropane	0.816	0.876	-7.4	101	0.00
27 T	cis-1,2-Dichloroethene	0.624	0.637	-2.1	98	0.00
28 T	Bromochloromethane	0.278	0.299	-7.6	95	-0.01
29 T	Tetrahydrofuran	0.281	0.267	5.0	88	0.00
30 C	Chloroform	0.995	1.018	-2.3#	100	0.00
31 T	Cyclohexane	0.907	0.945	-4.2	101	0.00
32 T	1,1,1-Trichloroethane	0.843	0.896	-6.3	99	0.00
33 S	1,2-Dichloroethane-d4	0.625	0.597	4.5	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.287	0.295	-2.8	96	0.00
36 T	1,1-Dichloropropene	0.440	0.473	-7.5	98	0.00
37 T	Ethyl Acetate	0.492	0.498	-1.2	91	0.00
38 T	Carbon Tetrachloride	0.428	0.465	-8.6	97	0.00
39 T	Methylcyclohexane	0.533	0.611	-14.6	102	0.00
40 TM	Benzene	1.298	1.404	-8.2	98	0.00
41 T	Methacrylonitrile	0.270	0.281	-4.1	96	0.00
42 TM	1,2-Dichloroethane	0.466	0.501	-7.5	98	0.00
43 T	Isopropyl Acetate	0.789	0.826	-4.7	94	0.00
44 TM	Trichloroethene	0.360	0.374	-3.9	99	0.00
45 C	1,2-Dichloropropane	0.331	0.354	-6.9#	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX100919\
 Data File : VX012929.D
 Acq On : 09 Oct 2019 09:33
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 10 01:56:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.220	0.240	-9.1	98	0.00
47 T	Bromodichloromethane	0.430	0.473	-10.0	97	0.00
48 T	Methyl methacrylate	0.384	0.398	-3.6	93	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	85	0.00
50 S	Toluene-d8	1.119	1.125	-0.5	95	0.00
51 T	4-Methyl-2-Pentanone	0.492	0.499	-1.4	91	0.00
52 CM	Toluene	0.831	0.895	-7.7#	100	0.00
53 T	t-1,3-Dichloropropene	0.476	0.525	-10.3	96	0.00
54 T	cis-1,3-Dichloropropene	0.529	0.588	-11.2	98	0.00
55 T	1,1,2-Trichloroethane	0.329	0.348	-5.8	97	0.00
56 T	Ethyl methacrylate	0.532	0.574	-7.9	95	0.00
57 T	1,3-Dichloropropane	0.569	0.598	-5.1	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.241	0.262	-8.7	95	0.00
59 T	2-Hexanone	0.383	0.405	-5.7	94	0.00
60 T	Dibromochloromethane	0.322	0.361	-12.1	95	0.00
61 T	1,2-Dibromoethane	0.346	0.370	-6.9	98	0.00
62 S	4-Bromofluorobenzene	0.435	0.441	-1.4	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.347	0.388	-11.8	103	0.00
65 PM	Chlorobenzene	0.993	1.043	-5.0	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.344	0.382	-11.0	97	0.00
67 C	Ethyl Benzene	1.764	1.928	-9.3#	98	0.00
68 T	m/p-Xylenes	0.659	0.727	-10.3	99	0.00
69 T	o-Xylene	0.643	0.699	-8.7	98	0.00
70 T	Styrene	1.092	1.214	-11.2	98	0.00
71 P	Bromoform	0.248	0.276	-11.3	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	3.676	3.991	-8.6	100	0.00
74 T	N-amyl acetate	1.583	1.689	-6.7	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.223	1.249	-2.1	97	0.00
76 T	1,2,3-Trichloropropane	1.109	1.045	5.8	99	0.00
77 T	Bromobenzene	0.866	0.902	-4.2	98	0.00
78 T	n-propylbenzene	4.045	4.493	-11.1	104	0.00
79 T	2-Chlorotoluene	2.542	2.665	-4.8	99	0.00
80 T	1,3,5-Trimethylbenzene	3.070	3.321	-8.2	99	0.00
81 T	trans-1,4-Dichloro-2-butene	0.392	0.410	-4.6	95	0.00
82 T	4-Chlorotoluene	2.914	3.120	-7.1	99	0.00
83 T	tert-Butylbenzene	2.994	3.165	-5.7	97	0.00
84 T	1,2,4-Trimethylbenzene	3.109	3.402	-9.4	99	0.00
85 T	sec-Butylbenzene	3.469	3.782	-9.0	98	0.00
86 T	p-Isopropyltoluene	3.217	3.591	-11.6	102	0.00
87 T	1,3-Dichlorobenzene	1.609	1.668	-3.7	98	0.00
88 T	1,4-Dichlorobenzene	1.629	1.702	-4.5	101	0.00
89 T	n-Butylbenzene	2.704	3.134	-15.9	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX100919\
Data File : VX012929.D
Acq On : 09 Oct 2019 09:33
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 10 01:56:52 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.521	0.589	-13.1	98	0.00
91 T	1,2-Dichlorobenzene	1.577	1.655	-4.9	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.287	0.297	-3.5	91	0.00
93 T	1,2,4-Trichlorobenzene	0.988	1.063	-7.6	98	0.00
94 T	Hexachlorobutadiene	0.456	0.494	-8.3	103	0.00
95 T	Naphthalene	3.425	3.644	-6.4	94	0.00
96 T	1,2,3-Trichlorobenzene	1.021	1.048	-2.6	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6