

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100920\
 Data File : VX018859.D
 Acq On : 09 Oct 2020 16:59
 Operator : JC/SP
 Sample : L4358-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW

Quant Time: Oct 10 06:44:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	223043	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	352867	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	330140	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	170344	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	162765	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
35) Dibromofluoromethane	5.46	113	124321	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
50) Toluene-d8	8.70	98	420320	48.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.36%	
62) 4-Bromofluorobenzene	11.12	95	160600	46.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.38%	
Target Compounds						
16) Acetone	2.42	43	7513	0.884	ug/l #	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100920\
Data File : VX018859.D
Acq On : 09 Oct 2020 16:59
Operator : JC/SP
Sample : L4358-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW

Quant Time: Oct 10 06:44:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

