

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100920\
 Data File : VX018864.D
 Acq On : 09 Oct 2020 18:53
 Operator : JC/SP
 Sample : L4260-17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Oct 10 06:55:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	211784	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	338478	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	321633	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	159665	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	159231	50.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.96%	
35) Dibromofluoromethane	5.47	113	118822	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
50) Toluene-d8	8.70	98	412217	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
62) 4-Bromofluorobenzene	11.12	95	152640	45.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.54%	
Target Compounds						
16) Acetone	2.42	43	10876	4.125	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100920\
Data File : VX018864.D
Acq On : 09 Oct 2020 18:53
Operator : JC/SP
Sample : L4260-17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Oct 10 06:55:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

