

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100920\
 Data File : VX018866.D
 Acq On : 09 Oct 2020 19:39
 Operator : JC/SP
 Sample : L4317-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 10 07:00:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	213478	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	348589	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	336115	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	176739	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	163733	52.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.00%	
35) Dibromofluoromethane	5.46	113	123310	49.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.28%	
50) Toluene-d8	8.70	98	426128	49.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.90%	
62) 4-Bromofluorobenzene	11.12	95	165062	48.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.12%	
Target Compounds						
20) Methylene Chloride	2.84	84	6064	3.177	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100920\
Data File : VX018866.D
Acq On : 09 Oct 2020 19:39
Operator : JC/SP
Sample : L4317-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Oct 10 07:00:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

