

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100920\
 Data File : VX018881.D
 Acq On : 10 Oct 2020 02:27
 Operator : JC/SP
 Sample : L4303-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FT-MW06S-20201002

Quant Time: Oct 10 07:36:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	200363	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	311436	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	311676	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	154005	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	152815	51.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.42%	
35) Dibromofluoromethane	5.46	113	113049	50.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.86%	
50) Toluene-d8	8.70	98	383260	49.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.56%	
62) 4-Bromofluorobenzene	11.12	95	147327	48.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.02%	
Target Compounds						
3) Chloromethane	1.31	50	2575	1.371	ug/l #	85
20) Methylene Chloride	2.84	84	1745	1.323	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100920\
Data File : VX018881.D
Acq On : 10 Oct 2020 02:27
Operator : JC/SP
Sample : L4303-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FT-MW06S-20201002

Quant Time: Oct 10 07:36:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

