

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100920\
 Data File : VX018882.D
 Acq On : 10 Oct 2020 02:50
 Operator : JC/SP
 Sample : L4303-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FT-MW06I-20201002

Quant Time: Oct 10 07:38:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	201773	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	322639	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	323644	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	162399	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	156054	52.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.88%	
35) Dibromofluoromethane	5.46	113	116454	50.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.30%	
50) Toluene-d8	8.70	98	403265	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
62) 4-Bromofluorobenzene	11.12	95	154401	48.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.14%	
Target Compounds						
3) Chloromethane	1.31	50	2983	1.577	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100920\
Data File : VX018882.D
Acq On : 10 Oct 2020 02:50
Operator : JC/SP
Sample : L4303-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FT-MW06I-20201002

Quant Time: Oct 10 07:38:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

