

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100920\
 Data File : VX018883.D
 Acq On : 10 Oct 2020 03:13
 Operator : JC/SP
 Sample : L4303-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FT-MW02S-20201002

Quant Time: Oct 10 07:40:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	199947	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	319208	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	316019	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	165385	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	152938	51.86	ug/l	0.00
Spiked Amount			Recovery	=	103.72%	
35) Dibromofluoromethane	5.47	113	113424	49.86	ug/l	0.00
Spiked Amount			Recovery	=	99.72%	
50) Toluene-d8	8.70	98	392101	49.69	ug/l	0.00
Spiked Amount			Recovery	=	99.38%	
62) 4-Bromofluorobenzene	11.12	95	157322	50.02	ug/l	0.00
Spiked Amount			Recovery	=	100.04%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
24) 1,1-Dichloroethane	3.67	63	12526	3.109	ug/l	97
27) cis-1,2-Dichloroethene	4.57	96	3528	1.483	ug/l	90
67) Ethyl Benzene	10.23	91	15986	1.388	ug/l	95
73) Isopropylbenzene	11.00	105	15432	1.426	ug/l	94
78) n-propylbenzene	11.34	91	31184	2.530	ug/l	100
83) tert-Butylbenzene	11.75	119	5810	0.627	ug/l	83
85) sec-Butylbenzene	11.93	105	14935	1.431	ug/l	100
87) 1,3-Dichlorobenzene	12.01	146	1390	0.255	ug/l	78
88) 1,4-Dichlorobenzene	12.08	146	14359	2.597	ug/l	89
89) n-Butylbenzene	12.37	91	7809	0.907	ug/l #	70
91) 1,2-Dichlorobenzene	12.38	146	45546	8.767	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100920\
Data File : VX018883.D
Acq On : 10 Oct 2020 03:13
Operator : JC/SP
Sample : L4303-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FT-MW02S-20201002

Quant Time: Oct 10 07:40:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

