

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100920\  
 Data File : VX018887.D  
 Acq On : 10 Oct 2020 04:43  
 Operator : JC/SP  
 Sample : L4255-13  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 44 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 FT-PZ-458I-20200928

Quant Time: Oct 10 07:48:38 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X100920W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Oct 10 06:09:09 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|-------|---------|----------|
| 1) Pentafluorobenzene       | 5.63  | 168  | 190949   | 50.00 | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.84  | 114  | 314325   | 50.00 | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 10.10 | 117  | 316877   | 50.00 | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.06 | 152  | 163245   | 50.00 | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |       |         |          |
| 33) 1,2-Dichloroethane-d4   | 6.03  | 65   | 153605   | 54.54 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 109.08% |          |
| 35) Dibromofluoromethane    | 5.46  | 113  | 115365   | 51.50 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 103.00% |          |
| 50) Toluene-d8              | 8.70  | 98   | 389482   | 50.12 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 100.24% |          |
| 62) 4-Bromofluorobenzene    | 11.12 | 95   | 152177   | 49.14 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 98.28%  |          |
| Target Compounds            |       |      |          |       |         |          |
| 27) cis-1,2-Dichloroethene  | 4.57  | 96   | 1758     | 0.774 | ug/l    | 82       |
| 44) Trichloroethene         | 7.19  | 130  | 13857    | 5.204 | ug/l    | 92       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100920\  
Data File : VX018887.D  
Acq On : 10 Oct 2020 04:43  
Operator : JC/SP  
Sample : L4255-13  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 44 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
FT-PZ-458I-20200928

Quant Time: Oct 10 07:48:38 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X100920W.M  
Quant Title : SW846 8260  
QLast Update : Sat Oct 10 06:09:09 2020  
Response via : Initial Calibration

