

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100920\
 Data File : VX018888.D
 Acq On : 10 Oct 2020 05:06
 Operator : JC/SP
 Sample : L4255-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FT-MW05S-20201001

Quant Time: Oct 10 07:50:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	193900	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	311499	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	296875	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	153345	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	149840	52.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.78%	
35) Dibromofluoromethane	5.47	113	111102	50.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.10%	
50) Toluene-d8	8.70	98	384831	49.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.96%	
62) 4-Bromofluorobenzene	11.12	95	143528	46.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.54%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100920\
Data File : VX018888.D
Acq On : 10 Oct 2020 05:06
Operator : JC/SP
Sample : L4255-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FT-MW05S-20201001

Quant Time: Oct 10 07:50:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

