

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX100920\
 Data File : VX018871.D
 Acq On : 09 Oct 2020 21:32
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 10 07:13:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	50.000	49.961	0.1	96	0.00
3 P	Chloromethane	50.000	50.722	-1.4	100	0.00
4 C	Vinyl Chloride	50.000	49.100	1.8#	97	0.00
5 T	Bromomethane	50.000	50.259	-0.5	102	0.00
6 T	Chloroethane	50.000	49.348	1.3	102	0.00
7 T	Trichlorofluoromethane	50.000	49.678	0.6	98	0.00
8 T	Diethyl Ether	50.000	50.530	-1.1	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.826	2.3	98	0.00
10 T	Methyl Iodide	50.000	36.968	26.1#	72	0.00
11 T	Tert butyl alcohol	250.000	239.364	4.3	94	0.00
12 CM	1,1-Dichloroethene	50.000	50.334	-0.7#	99	0.00
13 T	Acrolein	250.000	212.925	14.8	88	0.00
14 T	Allyl chloride	50.000	53.173	-6.3	106	0.00
15 T	Acrylonitrile	250.000	253.220	-1.3	99	0.00
16 T	Acetone	250.000	230.795	7.7	87	0.00
17 T	Carbon Disulfide	50.000	48.483	3.0	97	0.00
18 T	Methyl Acetate	50.000	49.515	1.0	103	0.00
19 T	Methyl tert-butyl Ether	50.000	51.175	-2.3	98	0.00
20 T	Methylene Chloride	50.000	51.161	-2.3	101	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.374	-0.7	101	0.00
22 T	Diisopropyl ether	50.000	50.693	-1.4	100	0.00
23 T	Vinyl Acetate	250.000	263.626	-5.5	98	0.00
24 P	1,1-Dichloroethane	50.000	47.616	4.8	99	0.00
25 T	2-Butanone	250.000	242.406	3.0	94	0.00
26 T	2,2-Dichloropropane	50.000	43.708	12.6	89	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.447	3.1	97	0.00
28 T	Bromochloromethane	50.000	47.552	4.9	98	0.00
29 T	Tetrahydrofuran	250.000	256.228	-2.5	99	0.00
30 C	Chloroform	50.000	48.162	3.7#	99	0.00
31 T	Cyclohexane	50.000	50.449	-0.9	98	0.00
32 T	1,1,1-Trichloroethane	50.000	48.000	4.0	98	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.461	1.1	102	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	97	0.00
35 S	Dibromofluoromethane	50.000	49.251	1.5	100	0.00
36 T	1,1-Dichloropropene	50.000	47.462	5.1	98	0.00
37 T	Ethyl Acetate	50.000	50.371	-0.7	97	0.00
38 T	Carbon Tetrachloride	50.000	49.310	1.4	97	0.00
39 T	Methylcyclohexane	50.000	50.481	-1.0	100	0.00
40 TM	Benzene	50.000	48.689	2.6	98	0.00
41 T	Methacrylonitrile	50.000	48.884	2.2	100	0.00
42 TM	1,2-Dichloroethane	50.000	47.290	5.4	99	0.00
43 T	Isopropyl Acetate	50.000	49.703	0.6	98	0.00
44 TM	Trichloroethene	50.000	47.995	4.0	99	0.00
45 C	1,2-Dichloropropane	50.000	48.165	3.7#	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX100920\
 Data File : VX018871.D
 Acq On : 09 Oct 2020 21:32
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 10 07:13:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
46 T	Dibromomethane	50.000	49.161	1.7	100	0.00
47 T	Bromodichloromethane	50.000	48.826	2.3	98	0.00
48 T	Methyl methacrylate	50.000	50.153	-0.3	97	0.00
49 T	1,4-Dioxane	1000.000	964.689	3.5	98	0.00
50 S	Toluene-d8	50.000	50.313	-0.6	100	0.00
51 T	4-Methyl-2-Pentanone	250.000	267.091	-6.8	98	0.00
52 CM	Toluene	50.000	50.038	-0.1#	98	0.00
53 T	t-1,3-Dichloropropene	50.000	49.667	0.7	96	0.00
54 T	cis-1,3-Dichloropropene	50.000	49.028	1.9	97	0.00
55 T	1,1,2-Trichloroethane	50.000	48.348	3.3	98	0.00
56 T	Ethyl methacrylate	50.000	49.129	1.7	100	0.00
57 T	1,3-Dichloropropane	50.000	50.179	-0.4	99	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	258.211	-3.3	97	0.00
59 T	2-Hexanone	250.000	266.487	-6.6	97	0.00
60 T	Dibromochloromethane	50.000	48.950	2.1	99	0.00
61 T	1,2-Dibromoethane	50.000	49.097	1.8	97	0.00
62 S	4-Bromofluorobenzene	50.000	53.103	-6.2	104	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	103	0.00
64 T	Tetrachloroethene	50.000	46.524	7.0	94	0.00
65 PM	Chlorobenzene	50.000	47.036	5.9	99	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	48.069	3.9	103	0.00
67 C	Ethyl Benzene	50.000	49.002	2.0#	97	0.00
68 T	m/p-Xylenes	100.000	97.836	2.2	97	0.00
69 T	o-Xylene	50.000	49.353	1.3	98	0.00
70 T	Styrene	50.000	50.870	-1.7	97	0.00
71 P	Bromoform	50.000	47.834	4.3	99	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	100	0.00
73 T	Isopropylbenzene	50.000	50.688	-1.4	97	0.00
74 T	N-amyl acetate	50.000	54.068	-8.1	99	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.257	3.5	101	0.00
76 T	1,2,3-Trichloropropane	50.000	46.772	6.5	97	0.00
77 T	Bromobenzene	50.000	49.478	1.0	98	0.00
78 T	n-propylbenzene	50.000	50.213	-0.4	96	0.00
79 T	2-Chlorotoluene	50.000	48.972	2.1	96	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.277	-2.6	96	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	46.175	7.7	92	0.00
82 T	4-Chlorotoluene	50.000	49.586	0.8	96	0.00
83 T	tert-Butylbenzene	50.000	50.280	-0.6	97	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.636	-1.3	96	0.00
85 T	sec-Butylbenzene	50.000	50.473	-0.9	96	0.00
86 T	p-Isopropyltoluene	50.000	51.340	-2.7	95	0.00
87 T	1,3-Dichlorobenzene	50.000	47.721	4.6	96	0.00
88 T	1,4-Dichlorobenzene	50.000	48.684	2.6	101	0.00
89 T	n-Butylbenzene	50.000	49.610	0.8	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX100920\
Data File : VX018871.D
Acq On : 09 Oct 2020 21:32
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 10 07:13:46 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	50.000	46.040	7.9	93	0.00
91 T	1,2-Dichlorobenzene	50.000	48.236	3.5	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.249	3.5	98	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.657	2.7	98	0.00
94 T	Hexachlorobutadiene	50.000	50.075	-0.2	105	0.00
95 T	Naphthalene	50.000	46.155	7.7	94	0.00
96 T	1,2,3-Trichlorobenzene	50.000	49.579	0.8	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6