

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX100920\
 Data File : VX018873.D
 Acq On : 09 Oct 2020 22:40
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 10 07:15:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	50.000	52.752	-5.5	96	0.00
3 P	Chloromethane	50.000	52.023	-4.0	97	0.00
4 C	Vinyl Chloride	50.000	53.673	-7.3#	100	0.00
5 T	Bromomethane	50.000	57.639	-15.3	111	0.00
6 T	Chloroethane	50.000	49.925	0.2	97	0.00
7 T	Trichlorofluoromethane	50.000	53.205	-6.4	99	0.00
8 T	Diethyl Ether	50.000	53.493	-7.0	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.404	-2.8	98	0.00
10 T	Methyl Iodide	50.000	44.690	10.6	84	0.00
11 T	Tert butyl alcohol	250.000	275.385	-10.2	102	0.00
12 CM	1,1-Dichloroethene	50.000	52.844	-5.7#	99	0.00
13 T	Acrolein	250.000	268.462	-7.4	105	0.00
14 T	Allyl chloride	50.000	52.531	-5.1	99	0.00
15 T	Acrylonitrile	250.000	271.080	-8.4	101	0.00
16 T	Acetone	250.000	243.414	2.6	87	0.00
17 T	Carbon Disulfide	50.000	50.633	-1.3	96	0.00
18 T	Methyl Acetate	50.000	53.123	-6.2	104	0.00
19 T	Methyl tert-butyl Ether	50.000	54.247	-8.5	99	0.00
20 T	Methylene Chloride	50.000	53.933	-7.9	101	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.791	-3.6	99	0.00
22 T	Diisopropyl ether	50.000	53.848	-7.7	101	0.00
23 T	Vinyl Acetate	250.000	279.556	-11.8	99	0.00
24 P	1,1-Dichloroethane	50.000	50.978	-2.0	100	0.00
25 T	2-Butanone	250.000	258.568	-3.4	95	0.00
26 T	2,2-Dichloropropane	50.000	44.739	10.5	86	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.802	-1.6	97	0.00
28 T	Bromochloromethane	50.000	50.013	-0.0	98	0.00
29 T	Tetrahydrofuran	250.000	277.768	-11.1	102	0.00
30 C	Chloroform	50.000	51.021	-2.0#	99	0.00
31 T	Cyclohexane	50.000	53.467	-6.9	98	0.00
32 T	1,1,1-Trichloroethane	50.000	50.758	-1.5	98	0.00
33 S	1,2-Dichloroethane-d4	50.000	51.273	-2.5	100	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	96	0.00
35 S	Dibromofluoromethane	50.000	49.534	0.9	99	0.00
36 T	1,1-Dichloropropene	50.000	47.789	4.4	97	0.00
37 T	Ethyl Acetate	50.000	53.074	-6.1	100	0.00
38 T	Carbon Tetrachloride	50.000	50.423	-0.8	98	0.00
39 T	Methylcyclohexane	50.000	50.395	-0.8	98	0.00
40 TM	Benzene	50.000	49.687	0.6	99	0.00
41 T	Methacrylonitrile	50.000	50.802	-1.6	102	0.00
42 TM	1,2-Dichloroethane	50.000	49.372	1.3	102	0.00
43 T	Isopropyl Acetate	50.000	51.436	-2.9	100	0.00
44 TM	Trichloroethene	50.000	48.545	2.9	98	0.00
45 C	1,2-Dichloropropane	50.000	49.332	1.3#	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX100920\
 Data File : VX018873.D
 Acq On : 09 Oct 2020 22:40
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 10 07:15:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
46 T	Dibromomethane	50.000	49.603	0.8	99	0.00
47 T	Bromodichloromethane	50.000	49.311	1.4	98	0.00
48 T	Methyl methacrylate	50.000	52.058	-4.1	99	0.00
49 T	1,4-Dioxane	1000.000	940.284	6.0	94	0.00
50 S	Toluene-d8	50.000	50.293	-0.6	99	0.00
51 T	4-Methyl-2-Pentanone	250.000	278.110	-11.2	101	0.00
52 CM	Toluene	50.000	51.288	-2.6#	99	0.00
53 T	t-1,3-Dichloropropene	50.000	50.421	-0.8	96	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.242	-0.5	98	0.00
55 T	1,1,2-Trichloroethane	50.000	49.703	0.6	99	0.00
56 T	Ethyl methacrylate	50.000	49.619	0.8	100	0.00
57 T	1,3-Dichloropropane	50.000	51.377	-2.8	99	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	269.776	-7.9	100	0.00
59 T	2-Hexanone	250.000	276.450	-10.6	99	0.00
60 T	Dibromochloromethane	50.000	48.747	2.5	97	0.00
61 T	1,2-Dibromoethane	50.000	51.083	-2.2	100	0.00
62 S	4-Bromofluorobenzene	50.000	51.483	-3.0	99	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	98	0.00
64 T	Tetrachloroethene	50.000	52.384	-4.8	101	0.00
65 PM	Chlorobenzene	50.000	49.942	0.1	100	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.838	0.3	102	0.00
67 C	Ethyl Benzene	50.000	50.963	-1.9#	96	0.00
68 T	m/p-Xylenes	100.000	102.155	-2.2	96	0.00
69 T	o-Xylene	50.000	51.813	-3.6	98	0.00
70 T	Styrene	50.000	53.209	-6.4	97	0.00
71 P	Bromoform	50.000	49.576	0.8	98	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	97	0.00
73 T	Isopropylbenzene	50.000	52.285	-4.6	97	0.00
74 T	N-amyl acetate	50.000	55.939	-11.9	100	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.392	1.2	101	0.00
76 T	1,2,3-Trichloropropane	50.000	48.426	3.1	98	0.00
77 T	Bromobenzene	50.000	49.652	0.7	96	0.00
78 T	n-propylbenzene	50.000	52.171	-4.3	98	0.00
79 T	2-Chlorotoluene	50.000	51.352	-2.7	98	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.273	-4.5	96	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.440	1.1	97	0.00
82 T	4-Chlorotoluene	50.000	51.191	-2.4	97	0.00
83 T	tert-Butylbenzene	50.000	51.544	-3.1	97	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.433	-6.9	99	0.00
85 T	sec-Butylbenzene	50.000	52.379	-4.8	97	0.00
86 T	p-Isopropyltoluene	50.000	52.295	-4.6	95	0.00
87 T	1,3-Dichlorobenzene	50.000	47.493	5.0	93	0.00
88 T	1,4-Dichlorobenzene	50.000	49.537	0.9	100	0.00
89 T	n-Butylbenzene	50.000	50.839	-1.7	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX100920\
Data File : VX018873.D
Acq On : 09 Oct 2020 22:40
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 10 07:15:59 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	50.000	48.185	3.6	96	0.00
91 T	1,2-Dichlorobenzene	50.000	48.883	2.2	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.821	2.4	97	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.230	1.5	97	0.00
94 T	Hexachlorobutadiene	50.000	48.133	3.7	98	0.00
95 T	Naphthalene	50.000	48.537	2.9	97	0.00
96 T	1,2,3-Trichlorobenzene	50.000	51.580	-3.2	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6