

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100920\
 Data File : VX018885.D
 Acq On : 10 Oct 2020 03:58
 Operator : JC/SP
 Sample : L4303-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FC-MW10S-20201002

Quant Time: Oct 10 07:44:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 10 06:09:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	203534	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	325845	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	324747	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	164599	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	156466	52.12	ug/l	0.00
Spiked Amount			Recovery	=	104.24%	
35) Dibromofluoromethane	5.47	113	116849	50.32	ug/l	0.00
Spiked Amount			Recovery	=	100.64%	
50) Toluene-d8	8.70	98	402603	49.98	ug/l	0.00
Spiked Amount			Recovery	=	99.96%	
62) 4-Bromofluorobenzene	11.12	95	153220	47.73	ug/l	0.00
Spiked Amount			Recovery	=	95.46%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100920\
Data File : VX018885.D
Acq On : 10 Oct 2020 03:58
Operator : JC/SP
Sample : L4303-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FC-MW10S-20201002

Quant Time: Oct 10 07:44:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 10 06:09:09 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 746

Delta R.T. 0.00 min

Lab File: VX018885.D

Acq: 10 Oct 2020 03:58

Instrument :

MSVOA_X

ClientSampleId :


FC-MW10S-20201002

Tot Ion:168 Resp: 203534

Ion Ratio Lower Upper




168 100

99 62.7 46.1 69.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 52.122 ug/l

RT: 6.03 min Scan# 811

Delta R.T. 0.00 min

Lab File: VX018885.D


Acq: 10 Oct 2020 03:58

Tgt Ion: 65 Resp: 156466

Ion Ratio Lower Upper



65 100


67 48.8 0.0 97.0



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.83 min Scan# 942

Delta R.T. -0.01 min

Lab File: VX018885.D

Acq: 10 Oct 2020 03:58

Instrument :

MSVOA_X

ClientSampleId :

FC-MW10S-20201002


Tot Ion:114 Resp: 325845

Ion Ratio Lower Upper

114 100

63 20.9 0.0 42.4




88 16.9 0.0 32.6



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 50.319 ug/l

RT: 5.47 min Scan# 719

Delta R.T. 0.00 min

Lab File: VX018885.D

Acq: 10 Oct 2020 03:58


Tgt Ion:113 Resp: 116849

Ion Ratio Lower Upper

113 100

111 101.3 80.1 120.1



192 19.9 16.0 24.0




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.981 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. 0.00 min

Lab File: VX018885.D

Acq: 10 Oct 2020 03:58

Instrument :

MSVOA_X

ClientSampleId :


FC-MW10S-20201002

Tot Ion: 98 Resp: 402603

Ion Ratio Lower Upper

98 100

100 66.1 53.1 79.7



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX

8.70

250000

200000



150000

100000

50000

0

Time--> 8.60 8.70 8.80



#62

4-Bromofluorobenzene

Concen: 47.726 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. 0.00 min

Lab File: VX018885.D

Acq: 10 Oct 2020 03:58


Tgt Ion: 95 Resp: 153220

Ion Ratio Lower Upper

95 100

174 81.5 0.0 160.0

176 78.0 0.0 161.2



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

11.12


150000


100000

50000

0

Time--> 11.00 11.10 11.20





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.10 min Scan# 1478

Delta R.T. 0.00 min

Lab File: VX018885.D

Acq: 10 Oct 2020 03:58

Instrument :

MSVOA_X

ClientSampleId :

FC-MW10S-20201002


Tot Ion:117 Resp: 324747

Ion Ratio Lower Upper

117 100

82 53.3 43.4 65.2




119 30.7 28.3 42.5



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): V

Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.06 min Scan# 1800

Delta R.T. 0.00 min

Lab File: VX018885.D

Acq: 10 Oct 2020 03:58


Tgt Ion:152 Resp: 164599

Ion Ratio Lower Upper

152 100

115 58.8 41.8 125.4

150 157.0 0.0 347.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

