


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX100920\
 Data File : VX018848.D
 Acq On : 09 Oct 2020 10:34
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Title : SW846 8260
 Last Update : Sat Oct 10 06:09:09 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	6616	PASS
75	95	30	60	51.5	17494	PASS
95	95	100	100	100.0	33981	PASS
96	95	5	9	6.4	2188	PASS
173	174	0.00	2	1.8	509	PASS
174	95	50	100	84.4	28672	PASS
175	174	5	9	7.3	2087	PASS
176	174	95	101	97.5	27946	PASS
177	176	5	9	6.0	1673	PASS