


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX100920\
 Data File : VX018872.D
 Acq On : 09 Oct 2020 21:55
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100920W.M
 Title : SW846 8260
 Last Update : Sat Oct 10 06:09:09 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.7	10001	PASS
75	95	30	60	53.2	28525	PASS
95	95	100	100	100.0	53605	PASS
96	95	5	9	7.0	3727	PASS
173	174	0.00	2	1.3	548	PASS
174	95	50	100	81.5	43696	PASS
175	174	5	9	7.5	3286	PASS
176	174	95	101	97.7	42701	PASS
177	176	5	9	6.8	2895	PASS