

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100923\
 Data File : VX038161.D
 Acq On : 09 Oct 2023 11:04
 Operator : JC/MD
 Sample : VX1009WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1009WBL01

Quant Time: Oct 10 00:51:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	123017	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	222375	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218747	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110701	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	90675	45.827	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.660%
35) Dibromofluoromethane	5.385	113	74455	46.466	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.940%
50) Toluene-d8	8.647	98	278961	45.560	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.120%#
62) 4-Bromofluorobenzene	11.079	95	104017	44.810	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100923\
Data File : VX038161.D
Acq On : 09 Oct 2023 11:04
Operator : JC/MD
Sample : VX1009WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1009WBL01

Quant Time: Oct 10 00:51:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

