

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100923\
 Data File : VX038171.D
 Acq On : 09 Oct 2023 15:00
 Operator : JC/MD
 Sample : 04810-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10A-GW-640-642

Quant Time: Oct 10 00:57:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	110758	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	197359	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	192227	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96159	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82063	46.065	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.120%
35) Dibromofluoromethane	5.391	113	64587	45.417	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.840%
50) Toluene-d8	8.647	98	243634	44.834	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.660%#
62) 4-Bromofluorobenzene	11.079	95	90560	43.958	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.920%
Target Compounds						
16) Acetone	2.374	43	1867	2.511	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100923\
Data File : VX038171.D
Acq On : 09 Oct 2023 15:00
Operator : JC/MD
Sample : 04810-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10A-GW-640-642

Quant Time: Oct 10 00:57:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

