

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100923\
 Data File : VX038177.D
 Acq On : 09 Oct 2023 17:19
 Operator : JC/MD
 Sample : 04724-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Oct 10 00:59:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	104112	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	184177	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182195	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91886	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77053	46.013	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.020%
35) Dibromofluoromethane	5.385	113	60199	45.361	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.720%
50) Toluene-d8	8.647	98	231400	45.630	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.260%#
62) 4-Bromofluorobenzene	11.079	95	85246	44.340	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100923\
Data File : VX038177.D
Acq On : 09 Oct 2023 17:19
Operator : JC/MD
Sample : 04724-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Oct 10 00:59:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

