

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100923\
 Data File : VX038178.D
 Acq On : 09 Oct 2023 17:42
 Operator : JC/MD
 Sample : 04713-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-1-10-13

Quant Time: Oct 10 00:59:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	107779	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	188306	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	189102	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96635	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	78813	45.463	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.920%
35) Dibromofluoromethane	5.385	113	62713	46.219	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.440%
50) Toluene-d8	8.647	98	236426	45.599	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.200%#
62) 4-Bromofluorobenzene	11.079	95	88736	45.143	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.280%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1983	2.741	ug/l	# 68
30) Chloroform	5.105	83	4496	1.833	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100923\
Data File : VX038178.D
Acq On : 09 Oct 2023 17:42
Operator : JC/MD
Sample : 04713-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-1-10-13

Quant Time: Oct 10 00:59:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

