

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100923\
 Data File : VX038180.D
 Acq On : 09 Oct 2023 18:29
 Operator : JC/MD
 Sample : 04713-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-1-37-40

Quant Time: Oct 10 01:00:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	106341	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	191541	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	188369	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97654	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79498	46.478	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	92.960%		
35) Dibromofluoromethane	5.385	113	62065	44.969	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	89.940%		
50) Toluene-d8	8.647	98	235452	44.644	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	89.280%#		
62) 4-Bromofluorobenzene	11.079	95	89119	44.572	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	89.140%		
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	14970	12.906	ug/l	98
12) 1,1-Dichloroethene	2.319	96	15270	15.695	ug/l	90
16) Acetone	2.373	43	3665	5.134	ug/l #	86
22) Diisopropyl ether	3.763	45	2976	0.816	ug/l #	74
24) 1,1-Dichloroethane	3.611	63	6884	3.110	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	5728	3.993	ug/l	81
30) Chloroform	5.105	83	2007	0.829	ug/l #	84
42) 1,2-Dichloroethane	6.098	62	3182	1.496	ug/l	93
44) Trichloroethene	7.129	130	16921	12.153	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100923\
Data File : VX038180.D
Acq On : 09 Oct 2023 18:29
Operator : JC/MD
Sample : 04713-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-1-37-40

Quant Time: Oct 10 01:00:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

