



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100923\
Data File : VX038158.D
Acq On : 09 Oct 2023 08:29
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Title : SW846 8260
Last Update : Thu Oct 05 15:27:13 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.4	10309	PASS
75	95	30	60	55.0	27813	PASS
95	95	100	100	100.0	50613	PASS
96	95	5	9	6.9	3513	PASS
173	174	0.00	2	0.9	364	PASS
174	95	50	100	76.7	38845	PASS
175	174	5	9	7.1	2776	PASS
176	174	95	101	97.3	37813	PASS
177	176	5	9	6.5	2469	PASS