

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100924\
 Data File : VX043334.D
 Acq On : 09 Oct 2024 13:12
 Operator : JC/MD
 Sample : PB163972ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB163972ZHE#03

Quant Time: Oct 10 01:30:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	87183	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	160993	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	134974	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58987	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	74094	52.038	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.080%
35) Dibromofluoromethane	5.385	113	53449	48.140	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.280%
50) Toluene-d8	8.647	98	191261	49.709	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.420%
62) 4-Bromofluorobenzene	11.079	95	67140	48.032	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.060%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	15105	25.403	ug/l	95
18) Methyl Acetate	2.709	43	3154	2.114	ug/l	97
20) Methylene Chloride	2.788	84	6351	5.793	ug/l	95
25) 2-Butanone	4.587	43	4396	5.392	ug/l	96
43) Isopropyl Acetate	6.342	43	86259	31.005	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100924\
Data File : VX043334.D
Acq On : 09 Oct 2024 13:12
Operator : JC/MD
Sample : PB163972ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB163972ZHE#03

Quant Time: Oct 10 01:30:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

