

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100924\
 Data File : VX043336.D
 Acq On : 09 Oct 2024 13:59
 Operator : JC/MD
 Sample : PB163972ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB163972ZHE#05

Quant Time: Oct 10 01:31:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	80888	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	151145	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126512	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55662	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69098	52.306	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.620%
35) Dibromofluoromethane	5.385	113	48977	46.987	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.980%
50) Toluene-d8	8.647	98	177359	49.099	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.200%
62) 4-Bromofluorobenzene	11.079	95	62158	47.366	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.740%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	15138	27.440	ug/l	96
18) Methyl Acetate	2.709	43	3221	2.327	ug/l	100
20) Methylene Chloride	2.788	84	6447	6.338	ug/l	98
25) 2-Butanone	4.580	43	4141	5.474	ug/l	96
43) Isopropyl Acetate	6.342	43	84879	32.496	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100924\
Data File : VX043336.D
Acq On : 09 Oct 2024 13:59
Operator : JC/MD
Sample : PB163972ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB163972ZHE#05

Quant Time: Oct 10 01:31:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

