

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100925\
 Data File : VX048076.D
 Acq On : 09 Oct 2025 11:29
 Operator : JC/MD
 Sample : Q3318-04 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 3436

Quant Time: Oct 10 01:19:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.531	168	117573	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.739	114	234000	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	239825	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	117315	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	92254	49.295	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.600%
35) Dibromofluoromethane	5.361	113	76915	49.011	ug/l	-0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.020%
50) Toluene-d8	8.628	98	262180	48.282	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.560%
62) 4-Bromofluorobenzene	11.061	95	111449	54.079	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.160%
Target Compounds						
16) Acetone	2.380	43	3214	3.950	ug/l #	85
59) 2-Hexanone	9.421	43	3447	2.150	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100925\
Data File : VX048076.D
Acq On : 09 Oct 2025 11:29
Operator : JC/MD
Sample : Q3318-04 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
3436

Quant Time: Oct 10 01:19:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

