

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100925\
 Data File : VX048088.D
 Acq On : 09 Oct 2025 15:42
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Oct 10 01:26:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	152762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	315518	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	318434	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	155266	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	124346	51.138	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.280%	
35) Dibromofluoromethane	5.373	113	105890	50.042	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.080%	
50) Toluene-d8	8.634	98	366109	50.002	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.000%	
62) 4-Bromofluorobenzene	11.061	95	154304	55.529	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 111.060%	
Target Compounds						Qvalue
64) Tetrachloroethene	9.256	164	55250	26.616	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100925\
Data File : VX048088.D
Acq On : 09 Oct 2025 15:42
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Oct 10 01:26:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

