

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100925\
 Data File : VX048090.D
 Acq On : 09 Oct 2025 16:24
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Oct 10 01:27:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	102037	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	211354	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	225749	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	115306	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	97785	60.206	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 120.420%#	
35) Dibromofluoromethane	5.373	113	78534	55.405	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.800%	
50) Toluene-d8	8.635	98	265497	54.131	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.260%	
62) 4-Bromofluorobenzene	11.061	95	115911	62.270	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 124.540%#	
Target Compounds						
64) Tetrachloroethene	9.256	164	37449	25.447	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100925\
Data File : VX048090.D
Acq On : 09 Oct 2025 16:24
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Oct 10 01:27:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

