

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100925\
 Data File : VX048091.D
 Acq On : 09 Oct 2025 16:44
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Oct 10 01:27:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	115160	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	244265	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	263898	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	134428	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	115943	63.251	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	126.500%#
35) Dibromofluoromethane	5.373	113	91850	56.068	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.140%
50) Toluene-d8	8.635	98	308899	54.495	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.980%
62) 4-Bromofluorobenzene	11.061	95	136952	63.661	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	127.320%#
Target Compounds						
						Qvalue
64) Tetrachloroethene	9.263	164	26210	15.235	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100925\
Data File : VX048091.D
Acq On : 09 Oct 2025 16:44
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Oct 10 01:27:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

