

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX100925\  
 Data File : VX048093.D  
 Acq On : 09 Oct 2025 17:25  
 Operator : JC/MD  
 Sample : IBLK  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 23 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 IBLK

Quant Time: Oct 10 01:28:35 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X091625W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 17 06:39:58 2025  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units       | Dev(Min)  |
|-----------------------------|--------|-------|----------|----------|-------------|-----------|
| -----                       |        |       |          |          |             |           |
| Internal Standards          |        |       |          |          |             |           |
| 1) Pentafluorobenzene       | 5.544  | 168   | 100134   | 50.000   | ug/l        | 0.00      |
| 34) 1,4-Difluorobenzene     | 6.751  | 114   | 208450   | 50.000   | ug/l        | 0.00      |
| 63) Chlorobenzene-d5        | 10.037 | 117   | 226846   | 50.000   | ug/l        | 0.00      |
| 72) 1,4-Dichlorobenzene-d4  | 12.006 | 152   | 115987   | 50.000   | ug/l        | 0.00      |
| System Monitoring Compounds |        |       |          |          |             |           |
| 33) 1,2-Dichloroethane-d4   | 5.940  | 65    | 99526    | 62.443   | ug/l        | 0.00      |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | = 124.880%# |           |
| 35) Dibromofluoromethane    | 5.379  | 113   | 77151    | 55.187   | ug/l        | 0.00      |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | = 110.380%  |           |
| 50) Toluene-d8              | 8.635  | 98    | 264132   | 54.603   | ug/l        | 0.00      |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | = 109.200%  |           |
| 62) 4-Bromofluorobenzene    | 11.067 | 95    | 116417   | 63.414   | ug/l        | 0.00      |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | = 126.820%# |           |
| Target Compounds            |        |       |          |          |             |           |
| 64) Tetrachloroethene       | 9.257  | 164   | 15591    | 10.543   | ug/l        | Qvalue 95 |
| -----                       |        |       |          |          |             |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX100925\  
Data File : VX048093.D  
Acq On : 09 Oct 2025 17:25  
Operator : JC/MD  
Sample : IBLK  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 23 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
IBLK

Quant Time: Oct 10 01:28:35 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X091625W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 17 06:39:58 2025  
Response via : Initial Calibration

