

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100925\
 Data File : VX048096.D
 Acq On : 09 Oct 2025 18:27
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Oct 10 01:29:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	114846	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	237992	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	253901	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	130525	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	110028	60.188	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	120.380%#
35) Dibromofluoromethane	5.379	113	89399	56.011	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.020%
50) Toluene-d8	8.635	98	300582	54.425	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.860%
62) 4-Bromofluorobenzene	11.061	95	131566	62.770	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	125.540%#
Target Compounds						
						Qvalue
64) Tetrachloroethene	9.262	164	8463	5.113	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100925\
Data File : VX048096.D
Acq On : 09 Oct 2025 18:27
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Oct 10 01:29:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

