

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100925\
 Data File : VX048097.D
 Acq On : 09 Oct 2025 18:47
 Operator : JC/MD
 Sample : Q3315-01MEDL 20X
 Misc : 3.10g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 314MEDL

Quant Time: Oct 10 01:30:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	100225	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	205661	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	217056	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	108316	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	89016	55.798	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.600%
35) Dibromofluoromethane	5.379	113	74520	54.028	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.060%
50) Toluene-d8	8.635	98	257302	53.913	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.820%
62) 4-Bromofluorobenzene	11.061	95	111277	61.436	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	122.880%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	26344	37.979	ug/l	96
52) Toluene	8.708	92	55534	14.725	ug/l	100
67) Ethyl Benzene	10.177	91	21563	2.534	ug/l	100
68) m/p-Xylenes	10.287	106	34138	10.780	ug/l	98
69) o-Xylene	10.628	106	9994	3.263	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100925\
Data File : VX048097.D
Acq On : 09 Oct 2025 18:47
Operator : JC/MD
Sample : Q3315-01MEDL 20X
Misc : 3.10g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
314MEDL

Quant Time: Oct 10 01:30:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

