

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX101018\
 Data File : VX005213.D
 Acq On : 10 Oct 2018 18:13
 Operator : JC/MD
 Sample : J5317-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-DUP04-20181004

Quant Time: Oct 11 07:15:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	167875	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	262213	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	255416	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	138859	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	128170	53.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.90%	
35) Dibromofluoromethane	5.51	113	109765	49.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.78%	
50) Toluene-d8	8.72	98	383906	51.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.92%	
62) 4-Bromofluorobenzene	11.14	95	133917	50.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.64%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.39	101	4233	2.30	ug/l	96
12) 1,1-Dichloroethene	2.37	96	7040	3.87	ug/l	88
16) Acetone	2.45	43	7126	5.09	ug/l	94
24) 1,1-Dichloroethane	3.70	63	19753	4.95	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	5334	2.40	ug/l	95
30) Chloroform	5.22	83	11490	3.05	ug/l	97
32) 1,1,1-Trichloroethane	5.49	97	2201	0.72	ug/l #	49
38) Carbon Tetrachloride	5.79	117	2267	0.71	ug/l	90
44) Trichloroethene	7.22	130	2356977	1010.06	ug/l	98
55) 1,1,2-Trichloroethane	9.21	97	3560	1.56	ug/l	92
95) Naphthalene	13.83	128	1684	0.86	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX101018\
Data File : VX005213.D
Acq On : 10 Oct 2018 18:13
Operator : JC/MD
Sample : J5317-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-DUP04-20181004

Quant Time: Oct 11 07:15:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

